


Pre-Pulse Inhibition of Startle Reflex (PPI)

User Guide

Rev. 2.0 - Dec. 2023



[ugobasile.com](https://www.ugobasile.com)



ugo basile®

YOUR COMPANION IN
DISCOVERY SINCE 1963

SAFETY CONSIDERATIONS

Although this instrument has been designed with international safety standards, it contains information, cautions and warnings which must be followed to ensure safe operation and to retain the instrument in safe conditions.

Service and adjustments should be carried out by qualified personnel, authorized by Ugo Basile organization.

Any adjustment, maintenance and repair of the powered instrument should be avoided as much as possible and, when inevitable, should be carried out by a skilled person who is aware of the hazard involved.

Capacitors inside the instrument may still be charged even if the instrument has been disconnected from its source of supply.



ugo basile®

TRANSFORMING IDEAS
INTO INSTRUMENTS

CE CONFORMITY STATEMENT

Manufacturer **UGO BASILE srl**
Address Via G. di Vittorio, 2 – 21036 Gemonio, VA, ITALY
Phone n. +39 0332 744574
Fax n. +39 0332 745488

We hereby declare that

Instrument: **Startle/PPI System**
Catalog number: **48163, 48263, 48463, 48162, 48262, 48462**

*is manufactured in compliance with the following European Union Directives
and relevant harmonized standards*

- *2014/35/UE relating to electrical equipment designed for use within certain voltage limits*
- *2014/30/UE relating to electromagnetic compatibility*
- *2011/65/UE and 2015/863/UE on the restriction of the use of certain hazardous substances in electrical and electronic equipment*

Account Manager

Mauro Uboldi

Nome / Name

December 2023

Date

Firma / Signature

Table of Contents

1.	Product features and general information	6
2.	What's in the box	7
3.	General	11
4.	Instrument description	12
5.	Installation	13
5.1.	Unpacking & preliminary check	13
5.2.	Safety instruction	13
5.3.	Intended Use	13
5.4.	Additional Safety Consideration.....	13
5.5.	Assembling the instrument	14
5.6.	Installing the Startle PPI software	16
5.7.	Starting the application.....	18
5.8.	Using the application.....	19
6.	Experiment Page	20
6.1.	New Experiment.....	21
6.2.	Load experiment	21
6.3.	Save experiment.....	21
6.4.	Print Experiment	21
6.5.	Add new treatment	22
6.6.	Remove Selected treatment.....	22
6.7.	Add New Stage	22
6.8.	Remove Selected stage	22
6.9.	Add new animals.....	23
6.10.	Remove Selected animal	23
6.11.	Trials table.....	23
7.	Protocol Page	25
7.1.	Protocol menu	25
7.1.1.	New File	25
7.1.2.	Load File	25
7.1.3.	Save File	25
7.1.4.	Save file as.....	26
7.1.5.	Check current.....	26
7.1.6.	Stop checking.....	26
7.1.7.	Print.....	26
7.2.	Protocol Blocks	26
7.3.	Comment.....	27
7.3.1.	Environment	27
7.3.2.	Wait/Pause	28
7.3.3.	Recording	28
7.3.4.	Loop.....	30
7.3.5.	Protocol End	30
7.3.6.	Stimuli	30
7.3.7.	Randomized stimuli	31
7.4.	Creating the protocol.....	32
7.5.	Real-world example	34

8.	Run Page	35
8.1.	Toggle Sensitivity Threshold	36
8.2.	Running a test	37
9.	View Page	38
9.1.	Experiment sessions	38
9.2.	Data grid area.....	38
9.3.	Plot chart area	39
9.4.	Stimuli exclusion.....	40
9.5.	Threshold adjustment	40
9.6.	Exporting data	41
9.6.1.	Raw data (XLSX only).....	41
9.6.2.	Startle data (XLSX only).....	41
9.7.	Statistical data	42
10.	Control Page	43
11.	Settings Page	44
11.1.	Working Database.....	44
11.2.	Firmware Update.....	44
12.	Maintenance	46
12.1.	Cleaning/disinfection.....	46
12.2.	Long Inactivity	46
12.3.	Customer Support.....	46
13.	Specification	47
14.	Bibliografy	48
15.	Related products	49

1. Product features and general information

One of the major challenges in Neuroscience is reproducing core symptoms of complex diseases, such as schizophrenia working with laboratory animals.

However, pre-pulse inhibition of the startle reflex represents an effective way to solve this challenge and has been used by scientists worldwide with great success, if one considers the more than 10,000 papers published using this methodology

Startle-PPI devices allow the quantification of the pre-pulse inhibition (PPI) of the startle reflex.

The startle magnitude elicited by an intense stimulus (auditory) is reduced when it is preceded by a weaker pre-pulse stimulus.

Impaired PPI is observed in schizophrenia (Braff et al. 2001, Swerdlow et al. 2008), Tourette's syndrome (Swerdlow et al. 2001), Huntington disease (Swerdlow et al 2005), Bipolar disorder (Perry et al. 2001).

This system support a maximum of 4 Startle PPI cages.

Available configurations are 1, 2, 3 or a maximum of 4 Startle-PPI cages

The Ugo Basile system is fully automated, both for the preparation of the procedure and for the analysis of data.

Main features include:

- Automatic detection of Startle Reflex in rodents
- Complete set up, ready to start your tests. Customizable to suit your needs. A typical system combines the following elements:
 - Dedicated Software
 - Up to 4 Isolation Cubicles that includes an infrared (I.R.) light, a loudspeaker and a noiseless fan, all conveniently positioned inside the sound attenuating box
 - Stimulating/Recording (S/R) Platform that includes the light (only for mouse version), speaker and detection system
- Multiple Animal restrainers to match animal dimension
- Flexible testing, easy to set up. Simply define trial number, sound and light stimulus and timing of experimental sequences (all fully randomizable):
 - Pulse
 - Pre-pulse
 - Inter-Pulse Interval
 - Inter-Stimulus Interval
 - TTL Outoputs to control external devices

2. What's in the box

Startle/PPI System for Mice, single cage system - SKU 48163

SKU	Q.ty	Description
46000-596	1	Isolation Cubicle for behavioral cages, inside dimensions 39(d)x47(w)x49(h)cm
46000-105	1	On-Board Controller, managing shock, light and sound. Complete with hub for videocamera connection and cable kit
46000-165	1	Sound Generator
46000-325	1	Dual visible/I.R. light
48003	1	Set of two Mouse Holders for Startle/PPI; small cage 40x90x40mm, large cage 45x90x45mm both compatible with 48006
48006	1	Stimulating/Recording Platform for Mouse
48163-150	1	Startle/PPI Link-Box
52010-323	1	USB stick, including instruction manual

Startle/PPI System for Mice, two-cage system - SKU 48263

SKU	Q.ty	Description
46000-596	2	Isolation Cubicle for behavioral cages, inside dimensions 39(d)x47(w)x49(h)cm
46000-105	2	On-Board Controller, managing shock, light and sound. Complete with hub for videocamera connection and cable kit
46000-165	2	Sound Generator
46000-325	2	Dual visible/I.R. light
48003	2	Set of two Mouse Holders for Startle/PPI; small cage 40x90x40mm, large cage 45x90x45mm both compatible with 48006
48006	2	Stimulating/Recording Platform for Mouse
48163-150	2	Startle/PPI Link-Box
52010-323	1	USB stick, including instruction manual

Startle/PPI System for Mice, four-cage system - SKU 48463

SKU	Q.ty	Description
46000-596	4	Isolation Cubicle for behavioral cages, inside dimensions 39(d)x47(w)x49(h)cm
46000-105	4	On-Board Controller, managing shock, light and sound. Complete with hub for videocamera connection and cable kit
46000-165	4	Sound Generator
46000-325	4	Dual visible/I.R. light
48003	4	Set of two Mouse Holders for Startle/PPI; small cage 40x90x40mm, large cage 45x90x45mm both compatible with 48006
48006	4	Stimulating/Recording Platform for Mouse
48163-150	4	Startle/PPI Link-Box
52010-323	1	USB stick, including instruction manual

Startle/PPI System for Rats, single cage system - SKU 48162

SKU	Q.ty	Description
46000-596	1	Isolation Cubicle for behavioral cages, inside dimensions 39(d)x47(w)x49(h)cm
46000-105	1	On-Board Controller, managing shock, light and sound. Complete with hub for videocamera connection and cable kit
46000-165	1	Sound Generator
46000-325	1	Dual visible/I.R. light
48002	1	Stimulating/Recording Platform for Rat
48163-150	1	Startle/PPI Link-Box
52010-323	1	USB stick, including instruction manual
One of the following restrainers is included in the System (depending on the animal weight communicated when placing the order)		
48003-360	-	Restrainer for rodents up to 100gr
48003-370	-	Restrainer for rodents up to 250gr
48003-390	-	Restrainer for rodents up to 450gr

Startle/PPI System for Rats, two-cage system - SKU 48262

SKU	Q.ty	Description
46000-596	2	Isolation Cubicle for behavioral cages, inside dimensions 39(d)x47(w)x49(h)cm
46000-105	2	On-Board Controller, managing shock, light and sound. Complete with hub for videocamera connection and cable kit
46000-165	2	Sound Generator
46000-325	2	Dual visible/I.R. light
48002	2	Stimulating/Recording Platform for Rat
48163-150	2	Startle/PPI Link-Box
52010-323	1	USB stick, including instruction manual
A set of 2 of the following restrainers is included in the System (depending on the animals weight communicated when placing the order)		
48003-360	-	Restrainer for rodents up to 100gr
48003-370	-	Restrainer for rodents up to 250gr
48003-390	-	Restrainer for rodents up to 450gr

Startle/PPI System for Rats, four-cage system - SKU 48462

SKU	Q.ty	Description
46000-596	4	Isolation Cubicle for behavioral cages, inside dimensions 39(d)x47(w)x49(h)cm
46000-105	4	On-Board Controller, managing shock, light and sound. Complete with hub for videocamera connection and cable kit
46000-165	4	Sound Generator
46000-325	4	Dual visible/I.R. light
48002	4	Stimulating/Recording Platform for Rat
48163-150	4	Startle/PPI Link-Box
52010-323	1	USB stick, including instruction manual
A set of 4 of the following restrainers is included in the System (depending on the animals weight communicated when placing the order)		
48003-360	-	Restrainer for rodents up to 100gr
48003-370	-	Restrainer for rodents up to 250gr
48003-390	-	Restrainer for rodents up to 450gr

Optional items ordering information

SKU	Description
48000-200	4-Channel Air-Puff Dispenser, for Startle/PPI System. It requires connection to a cylinder or a gas line, with reducer. Each Startle/PPI unit requires in addition one kit 48000-204 including delivery system to stimulate animal with air.
48000-204	Air-puff kit, for each Startle/PPI Unit, to be attached to dispenser 48000-200. It includes all accessories for air delivery, except the animal cage (48003, two cage kit)

Available spare parts

SKU	Description
48002	Stimulating/Recording Platform for Rat
48006	Stimulating/Recording Platform for Mouse
48003	Set of two Mouse Holders for Startle/PPI; small cage 40x90x40mm, large cage 45x90x45mm both compatible with 48006
48003-360	Restrainer for rodents up to 100gr
48003-370	Restrainer for rodents up to 250gr
48003-390	Restrainer for rodents up to 450gr

3. General

Startle Prepulse Inhibition (PPI) is an experimental paradigm commonly used in rodents, such as mice and rats, to study sensorimotor gating, which refers to the brain's ability to filter out irrelevant sensory stimuli and prevent them from overwhelming the processing of more important stimuli.

Sensorimotor gating refers to the ability of a sensory event to suppress or decrease a motor response, in this case, a startle response

The Startle PPI paradigm has been widely used in neuroscience research to study sensorimotor gating deficits in different animal models, including genetic mouse models and those induced by pharmacological agents. It has also been employed to investigate the neural mechanisms underlying sensorimotor gating and to evaluate the effects of potential therapeutic interventions for disorders like schizophrenia, where PPI deficits are often observed.

A typical experiment includes a startle baseline measurement as a response to a loud sound (more rarely to a light or air-puff). A pre-pulse stimulus is delivered shortly before the loud stimulus and is supposed to decrease the startle reaction in healthy animals. Normally the time between trials is randomized to ensure that the animal response is not influenced by previous stimuli.

Researchers can manipulate various experimental parameters to investigate the factors influencing PPI, such as the intensity of the prepulse, the interval between the prepulse and startle stimulus (known as the prepulse-to-pulse interval or IPI Inter-Pulse-Interval), and the genetic or pharmacological manipulations of the animals.

4. Instrument description

The Ugo Basile Startle-PPI system is based on the light/sound pre-pulse and pulse stimulations, all embedded in a frame in which the animal cage and the force sensors are placed, together with the stimuli.

All is then positioned within an isolation cubicle to avoid external interferences and cross talk among cages, if more than one is used at the same time.

Currently, we support up to 4 cages connected to 1 PC.

Settings, protocol configuration, experiment run, management and results, are all controlled and displayed by a specific software.

Pulses and pre-pulses consist of white noise and frequency modulated sounds, lights or air-puffs.

The animal is restrained in a Plexiglass tube sitting on a force sensor. The dedicated software analyses data automatically, as startle response amplitude and %PPI.

The various procedures expose the animals to "Trials" of different nature:

- Pulse alone trial: only the strong stimulus is presented
- Pre-pulse alone trial: only the weak stimulus is presented
- Pre-pulse-pulse trial: the strong stimulus is preceded by the weak (pre-pulse) stimulus
- No-stim trials: no stimulus is delivered but the response of the animal is recorded to serve as a baseline

Experimental parameters can all be adjusted, linked, spaced in time, etc., to virtually build in a seamless and user-friendly manner virtually any type of startle and startle/PPI protocols.

5. Installation

5.1. Unpacking & preliminary check

Check the contents of the shipment for completeness, check list to hand, and visually inspect the instrument as soon you take it out the packaging.

If the instrument is damaged, immediately inform the shipping agent or carrier and notifying us by email at support@ugobasile.com.

Inspect the instrument for damages such as scratches, broken or loose parts. If after having tested it, the instrument fails to meet rated performances, contact our company immediately via email at service@ugobasile.com.

Notes on the User Guide

This User Guide can be found on the provided USB Key. We recommend to read thoroughly this manual, as it contains essential informations for the correct installation and operation of the instrument.

Please save the manual in a safe place, ready to be consulted by qualified personnel who use the instrument. Print it only if necessary.

Our User Guide are available as free download on our web site. For any additional information and/or assistance, you are welcome to contact our Service Department at service@ugobasile.com, specifying the serial number of the instrument.

5.2. Safety instruction

To ensure safe operation DO NOT attempt to open or perform any service work before having contacted Ugo Basile support team.

Pay attention while this device can emit a very loud sound at low and high frequencies that can damage human and animal auricle; always close the sound isolation cubicle door before running the test.

5.3. Intended Use

The Startle PPI is intended for **Investigation use on laboratory animal only. DO NOT USE ON HUMANS.**

5.4. Additional Safety Consideration

1. Use original accessories and spare parts only.
2. Immediately disconnect and replace damaged main cord.
3. Do not obstruct access to the power module.
4. Do not operate in hazardous environment or outside prescribed environmental limitation.
5. Do not spray any liquid on the connectors or on the electronic unit.

Ugo Basile cannot in any way and form be held responsible for damage caused to things and people and warranty will be void, due to:

- Incorrect electrical supply.
- Incorrect installation procedure.
- Incorrect or improper use or, in any case, not in accordance with the purpose for which the instrument has been designed and the warnings stated in the instruction manual supplied with the instrument.
- Replacement of original component, accessories or parts with others not made by Ugo Basile.
- Servicing carried out by unauthorized personnel.

5.5. Assembling the instrument


Place the isolation cubicle on a flat surface (usually on a desk) pay attention to leave a minimum of 15 cm space around the rear, the left and the right side of it, to avoid air obstruction.

The surface must be stable and should not propagate, from the table or the floor.



Place the Startle PPI cage inside the isolation cubicle in the centre of the floor.

Make the correct electric and pneumatic connection according to the following diagram:

Startle PPI Connection diagram



Cubicle inside



5.6. Installing the Startle PPI software

Startle PPI software system requirements:

Feature	Minimum	Recommended
CPU	Intel i7	Intel i9
Memory (RAM)	8 GB	16 GB
Storage (free space)	1 GB	10 GB
Monitor	1024 x 768	1920 x 1080
Operative system	Microsoft Windows 10 64bit	Microsoft Windows 11 64bit
Keyboard	Yes	Yes
Mouse	Yes	Yes
USB port	1 free (USB 2.0)	1 free (USB 2.0) for each cage


You will find the software installation kit inside the provided USB memory key.

The installation package contains two files:




Double click the setup.exe application to begin the install process. If an older version of the application is already installed it is overwritten with the new version. Data is not deleted by the setup process.

After starting the setup process this window appears:




Click Next




Here you can change the installation path. If more users share the same machine with different profiles, it is suggested to choose “Everyone”.

Click Next



Click “Next” again to start installation process



When installation process ends, the following message is shown



A shortcut for starting the application is placed automatically in the Start menu and on the Desktop.

5.7. Starting the application

Before starting the application, ensure that the USB interface cable is connected to the available cubicles.

To start the application just double click the shortcut on the desktop. A splash screen appears for a few seconds while the application is loading.

STARTLE PPI 2.0

Pre-pulse inhibition software for mice and rats

This software has to be used only with the Ugo Basile Startle PPI hardware. Any other use of this software is prohibited.

2022 Ugo Basile SRL Italy, all right reserved.



Finalizing...



The program, once started, automatically creates a “StartlePPI” folder in your “Documents” folder (if not already present).

In this folder other subfolders are created, containing the following files and folders:

CalData	18/12/2023 10:25	Cartella di file
DB	07/12/2023 15:59	Cartella di file
Experiments	24/10/2023 11:46	Cartella di file
Export	13/11/2023 11:19	Cartella di file
RESULTS	21/03/2022 17:46	Cartella di file
Scripts	22/11/2023 13:57	Cartella di file
Updates	06/12/2023 08:52	Cartella di file
StartlePPI.config	07/12/2023 15:59	BDS.config

DB: contains the program database with all data resulting from the experiments done. The default file is named “**db.sqlite**”. If the file is deleted, a new empty file is created at the next program launch. (See paragraph “[11.2. Firmware Update](#)”)

EXPERIMENTS: this folder contains all the created experiments in XML format

SCRIPTS: this folder contains the protocol scripts and the file extension is .PPI.

EXPORT: this folder is used as the destination when exporting the experiments in CSV/XLSX format.

UPDATES: in this folder it is possible to eventually store firmware update files.

5.8. Using the application

General information

The application pages are identified by tabs:



The currently selected page is highlighted.

To close the application, click the icon in the top right corner.

On the bottom bar, the footer, you will find the current date and time and the program version.

6. Experiment Page

UB Startle PPI [db.sqlite]

EXPERIMENT PROTOCOL RUN VIEW CONTROL **SETTINGS**

New Load Save Export Print Add new Remove selected Add new Remove selected Add new Remove selected
Experiment Treatments Stages Animals

TREATMENTS

ID	Name	Animals
1	Treatment 1	0

STAGES

ID	Name	# of Trials
1	Stage 1	1

ANIMALS

ID	AnimalID	Gender	Weight	Color	Assign
----	----------	--------	--------	-------	--------

Experiment name: _____
Preview by # of cubicles: 1 Order by: Stage

TRIALS

Run	Animal ID	Color	Gender	Weight	Treatment	Stage	Trial
-----	-----------	-------	--------	--------	-----------	-------	-------

No communication port found. Please check the hardware or contact Ugo Basile for assistance.

18/12/2023 15:45:49 Version 2.0.21.1

This is the screen where you can define the information of the treatments, stages and animals involved in the experiment.

First of all it is necessary to define one or more treatments.

Then you can define one or more stages.

Then you can add the animals. Animals can be added in groups. Each animal can be assigned to one treatment only.

Adding treatments, stages (along with the number of trials for each stage) and animals, the right table is automatically populated.


The top menu offers these buttons:

EXPERIMENT PROTOCOL RUN VIEW CONTROL


New Load Save Export Print Add new Remove selected Add new Remove selected Add new Remove selected
Experiment Treatments Stages Animals

6.1. New Experiment

It allows to clear all current data and start entering new data for an experiment. A name for the experiment is necessary. A warning appears if unsaved data is present:



Then a dialog window opens, letting entering a name for the experiment.



Type in a name for the experiment and hit "OK". The new experiment is now ready to be filled with treatments, stages and animals. One treatment and one stage are created by default.

6.2. Load experiment

This button is used to load a previously saved experiment.

6.3. Save experiment

This button is used to save the current experiment. If you want to create a new experiment, please give a different name to the file to prevent overwriting the existing one.

6.4. Print Experiment

This button prints the current "Trials" table (the one in the right table of this page).

6.5. Add new treatment

This button adds a treatment to the treatments table. Treatments are automatically named "Treatment 1, Treatment 2 and so on. You can rename any treatment clicking the name in the treatments table:

Treatments (1)

ID	Name	Animals
1	<input type="text" value="My New Treatment Name"/>	1

6.6. Remove Selected treatment

It allows to remove the currently selected treatment. Only treatments with no assigned animals can be deleted.

6.7. Add New Stage

This button adds a stage to the stage table. Stages are automatically named "Stage 1, Stage 2 and so on. You can rename any stage clicking the name in the Stages table, same as explained before for the treatments. Moreover, here you can define how many times a stage must be repeated defining the value in the "# of Trials" column.

Stages (1)

ID	Name	# of Trials
1	Stage 1	<input type="text" value="3"/>

6.8. Remove Selected stage

It allows to remove the currently selected stage.

6.9. Add new animals

When this button is pressed, a dialog window opens, letting you to enter one or more animals:

Animals can be assigned to one only treatment. Animals are automatically named as Animal 1, Animal 2, and so on. Then, in the animals table, you can assign the desired Animal ID that uniquely identifies the animal.

6.10. Remove Selected animal

It allows to remove the currently selected animal.

6.11. Trials table

When any change is done to the treatments, stages and animals, it is automatically reflected to the trials table.

Experiment name: **temporary**

Preview by # of cubicles: 1 Order by: Stage

TRIALS (10)

Run	Animal ID	Color	Gender	Weight	Treatment	Stage	Trial
1	IT123	White	Male	30	Saline	Stage 1	1
2	ID345	White	Male	30	Saline	Stage 1	1
3	Animal 5	White	Male	30	Saline	Stage 1	1
4	Animal 3	White	Male	30	Drug	Stage 1	1
5	Animal 4	White	Male	30	Drug	Stage 1	1
6	IT123	White	Male	30	Saline	Stage 1	2
7	ID345	White	Male	30	Saline	Stage 1	2
8	Animal 5	White	Male	30	Saline	Stage 1	2
9	Animal 3	White	Male	30	Drug	Stage 1	2
10	Animal 4	White	Male	30	Drug	Stage 1	2
11	IT123	White	Male	30	Saline	Stage 2	1
12	ID345	White	Male	30	Saline	Stage 2	1
13	Animal 5	White	Male	30	Saline	Stage 2	1
14	Animal 3	White	Male	30	Drug	Stage 2	1
15	Animal 4	White	Male	30	Drug	Stage 2	1


In this example there are two treatments (named "Treated" and "Untreated") and 2 stages of two trials each. We Added 4 animals, that are assigned two to each treatment.

The first column, named "Run" indicates the experiment subsequent run, as it depends on the number of animals being tested concurrently (so it depends on the number of cubicles you will use). The number of cubicles can be selected in the "Preview by # of cubicles" dropdown list:

Preview by # of cubicle: 1 ▲

You can choose from 1 to 4 cubicles.


Moreover you can choose in which order you wish to test the animals (by treatment, by stage of by animal):




Regardless the chosen distribution mode and the number of cubicles used, in each run are assigned only animals that belong to the same treatment, stage and trial. So, in this example, no more than two animals can be tested at the same time.

7. Protocol Page

The protocol page allows to create a script for the experiment, where you define the environment characteristics and the stimuli sequence.



7.1. Protocol menu



7.1.1. New File

This button allows to create a new protocol

7.1.2. Load File

Allows to load a previously saved protocol

7.1.3. Save File

Saves the current protocol with the same name

7.1.4. Save file as...

Lets you to save the current protocol giving a new name.

7.1.5. Check current

Checks the current protocol for errors or wrong values

7.1.6. Stop checking


Stops the current check

7.1.7. Print

Prints out the current protocol.

7.2. Protocol Blocks

On the left side the block categories are listed and hereafter described. The blocks can be dragged and dropped after (or before) a block already present in the main area.




7.3. Comment



The **Comment block (//)** is used just as a remark to describe anything you want in the protocol. It does not make any action.

The **Step Name Block** lets you to enter a description that is then added to all the following stimuli, and it can be used to filter out data. For example, you can give names such as "Acclimation basal phase", "Pseudo random phase", "Verification basal phase" and so on.

7.3.1. Environment



White noise start lets you start a background white noise in the cubicle, defining the dB level from 45 to 120dB.

White noise stop disables the white noise output.

Fan Start starts the cubicle fan.

Fan Stop stops the cubicle fan.

Visible Light Start turns the visible light on defining the brightness from 10 to 100%.

Visible Light Stop turns the visible light off.

IR Light Start turns the IR light on defining the brightness from 10 to 100%.


IR Light Stop turns the IR light off.

Aux (Out 2) Start turns the auxiliary output on.

Aux (Out 2) Stop turns the auxiliary output off.

Stop All turns off all the environment outputs off.

7.3.2. Wait/Pause




Wait Exact adds a fixed delay. The value is in milliseconds, so a value of 10000ms corresponds to 10 seconds

Wait Random creates a random delay between the given limits

Wait random values peeks a delay between 5 given values

7.3.3. Recording



Recording Start: this block is used to define that the following stimuli will be recorded. Values for this block are described later.

Recording Stop: the following stimuli are not recorded.

The **START RECORDING** block allows to set the following values:

- **Capture window:** it indicates the time window that is recorded after the end of the stimulus.
- **Peak to peak:** it is the maximum time length of the startle reaction, from peak to peak. If the peaks time distance is greater than this value, the reaction will not be marked as a detection.
- **Pre-recording:** it is the time period that it is recorder prior to firing the stimulus. This can be useful to verify weather the animal was already moving in the cage before the stimulus is fired.
- **Detection time limit:** the startle reaction is searched from the stimulus start (or the beginning of the second pulse in the prepulse-pulse stimuli) until this time after the end of the stimulus.

The total time recorded for a stimulus is calculated with the sum of:

- Pre-recording time
- Pre-pulse duration (if it is a Prepulse-pulse stimulus)
- IPI, aka Inter-Pulse-Interval (if it is a Prepulse-pulse stimulus)
- Pulse duration
- Detection window

The startle event search is done:

- Pulse or Prepulse Alone stimuli: from the start of the pulse till the end of the pulse plus the detection time limit
- Prepulse-Pulse stimuli: from the start of the pulse till the end of the pulse plus the detection time limit

During

- the pre-recording
- the Prepulse and the IPI in the Prepulse-Pulse stimuli
- and after the Detection time limit

Any animal movement is recorded but not marked as a detection.

7.3.4. Loop

Repetition Loop



Repeat: you can define a loop to repeat the same sequence a number of times.


7.3.5. Protocol End

Protocol



End Trial: this block must be set at the end of the protocol.

7.3.6. Stimuli



NO STIMULUS is used for auditing the animal movements when no any stimulus is fired. The stimulus reaction is recorded in the same way of the other stimuli.

PREPULSE ALONE and **PULSE ALONE** are practically the same, they are just stimuli alone. You can set the duration and activate, for the given duration, light, air puff, auxiliary out, white noise (with dB amplitude) and sound (with dB amplitude and frequency): you can set one or more of these types of stimuli.

PREPULSE-PULSE is a combination where a sequence of two stimuli is fired, a pre-pulse and a pulse. IPI duration is the interval that lasts between the pre-pulse and the pulse and a random value between the two limits is used. Just put the same value for IPI Min and Max to set a constant interval.

7.3.7. Randomized stimuli

Randomization

RANDOMIZE: Delay between events from 10000 ms to 30000 ms

RANDOMIZED NO STIMULUS Duration: 40 ms
Occurrences: 1

RANDOMIZED PREPULSE ALONE Duration: 40 ms
Light: Air Puff: Aux (Out1):
White Noise: Amplitude: 65 dB
Sound: Amplitude: 65 dB Frequency: 2000 Hz
Occurrences: 1

RANDOMIZED PULSE ALONE Duration: 40 ms
Light: Air Puff: Aux (Out1):
White Noise: Amplitude: 120 dB
Sound: Amplitude: 120 dB Frequency: 2000 Hz
Occurrences: 1

RANDOMIZED PREPULSE-PULSE PrePulse Duration: 40 ms Light: AirPuff:
Aux (Out1):
WhiteNoise: Amplitude: 65 dB
Sound: Amplitude: 65 dB Frequency: 2000 Hz
IPI duration Min: 100 Max: 100 ms
Pulse Duration: 40 ms Light: AirPuff: Aux (Out1):
WhiteNoise: Amplitude: 120 dB
Sound: Amplitude: 120 dB Frequency: 2000 Hz
Occurrences: 1

RANDOMIZE: in this block you can put only the following “RANDOMIZED” stimuli and for each of them you must define the quantity of the stimulus to be randomly fired in the block. The delay between the stimuli is randomly chosen between the given limits in the RANDOMIZE block (in red).

If you need a protocol where the same stimuli set (for example 8 different stimuli) must be performed in a pseudo random sequence for example 10 times, you can create a randomize block with the set of 8 different stimuli with a quantity of 1 each and then put this block inside a REPEAT 10 times block (See Paragraph [“7.5. Real-world example” a pagina 34](#)).

Details for the stimuli blocks are exactly the same ones described in the previous section “Stimuli”. The only difference is that the randomized stimuli blocks have the “Occurrences” field, to define the quantity of the stimulus to be randomly fired in the block.

Warning: Light stimulus is available only for mouse platform.

7.4. Creating the protocol



The protocol script is shown in the main area of the screen. The first block is the purple block called "START".

The next blocks must be attached to this block. To attach a block, you need to drag the desired block and drop it under the previous block, when a green line appears:



Click the left side of the desired block and keep the left mouse button pressed



Keeping the left button pressed, drag the block.

As you can see, while dragging the block, on the left side a wastebasket icon appears, and you can release there the block to delete it.

Normally, drag it...




Drag it....



Until you reach the target position, when a green line appears under the preceding block.



While the green line is lit, release the mouse button and the block is automatically connected to the preceding block.



Now you can start to type in or edit the block properties...



In this example it is a comment block, so you can type in the comment



All set.

You can proceed to add the next blocks until you complete your protocol.

7.5. Real-world example

Starting block

Comment: set environment

Start white noise at 70dB

Start Fan

Wait 10 seconds

Start Recording

Basal phase

Repeat 5 times a pulse alone stimulus at 120dB of white noise

With a delay of 5s between them

Start the test

Repeat 10 times

a randomized block of 8 different stimuli with a random delay between 10 to 30s.

In this example we set:

- Prepulse-Pulse WN 73dB and Pulse WN 120dB
- Prepulse-Pulse WN 76dB and Pulse WN 120dB
- Prepulse-Pulse WN 82dB and Pulse WN 120dB
- Pulse Alone WN 120dB
- Prepulse Alone WN 73dB
- Prepulse Alone WN 76dB
- Prepulse Alone WN 82dB
- No Stimulus

One occurrence each. These 8 stimuli will be fired in a random sequence, so every one loop of the 10 set in the repeat block will produce different sequences of the eight stimuli.

Last phase

Repeat 5 times

A Pulse alone stimulus of 120dB White noise

with a random delay between them from 10 to 20s

Stop the recording

Stop the environment outputs

End the protocol

8. Run Page

This page is used for executing the experiment in the up-to-four connected cubicles.

UB Startle PPI [db.sqlite]

EXPERIMENT PROTOCOL **RUN** VIEW CONTROL SETTINGS

Toggle Sensitivity Threshold Load Protocol Run Test Stop Test Load Experiment Previous Next

Controller Experiment

Firmware	S/N					
7.27	FD/CYRHD					
Animal Id	Color	Weight	Gender	Treatment	Stage	Trial

Firmware	S/N					
N.C.	N.A.					
Animal Id	Color	Weight	Gender	Treatment	Stage	Trial

Firmware	S/N					
N.C.	N.A.					
Animal Id	Color	Weight	Gender	Treatment	Stage	Trial

Firmware	S/N					
N.C.	N.A.					
Animal Id	Color	Weight	Gender	Treatment	Stage	Trial

■ No stimulus ■ Pre-Pulse ■ Pulse | Detection

Selected Protocol:


Selected Experiment:

Elapsed time: 00:00 (remaining: from 00:00 to 00:00)

18/12/2023 15:45:49 Version 2.0.21.1

To the right of each connected cage, the firmware version and the serial number of the cubicle interface is shown.

8.1. Toggle Sensitivity Threshold



When activating the sensitivity threshold, you can set the upper and lower limits of the “noise” animal movement area. Keep the red line as narrow as possible. Movements whose peaks fall within the red area are not marked as a detection.

While the sensitivity toggle is activated, you can fire a 120dB white-noise pulse (Fire WN button) or a 3KHz sound pulse (Fire SND button), in order to check that the sound air movement doesn't affect the platform.

Please be aware that an empty restrainer placed on the cage can resonate as it acts as a resonating chamber. When the animal is placed in the restrainer this resonance is mostly attenuated.

When you are satisfied with the setting, press again the toggle sensitivity button to return to normal mode.

In any case, this threshold settings are provisional. You can change these thresholds later on (when the recording is completed).

8.2. Running a test

Before running a protocol, you must select it. Click “Load Protocol” and select the one you want to run.


Moreover, you can select the experiment, that is the list of animals to be tested, clicking “Load experiment”.

Animals are distributed to the active cages sequentially, following the sequence chosen in the EXPERIMENT page. You can change the animals pressing the previous-next buttons. The chosen animal is shown in the right area.

When you are ready, press the Run Test button to start it. A timer indicates the elapsed time and the minimum/maximum remaining time (depending on the random delays set in the protocol).

Every time you run a test, a new session is created and named “PPI_” followed by date and time.

In each connected cage’s row, the result of each stimulus is drawn:



The position of the pulses is highlighted using different colors for prepulse, pulse and no stimulus. This example shows a prepulse-pulse stimulus. As you can see in the example there is a 200ms (0.2s) pre recording time, the stimulus made of a 20ms prepulse, a 100ms inter stimuli delay and a 20ms pulse, then a 500ms capture window is recorded. The total recorded window is 840ms.

In the bottom right block the log of the executed operations is shown:

```
| Delay for 3000ms
Loop completed
STEP NAME: Phase B
Repeat 10 times...
| Start randomized block...
| | Randomized remaining stimuli in this block: 8
| | -> Randomized PREPULSE ALONE Stimulus, total duration 20ms [76dB]
| | -> ISI delay 5253ms (from 3000 to 6000)
```

The leading “|” (pipe) character is used to make evident the nested loops. Each “kind” of execution has its own color.

9. View Page

This page is used to analyze the result data of the experiments.

The screenshot shows the 'VIEW' tab of the software. It features a top menu with 'EXPERIMENT', 'PROTOCOL', 'RUN', 'VIEW', 'CONTROL', and 'SETTINGS'. Below the menu are icons for 'Export Selected', 'Export All', 'Export Select Fields', 'Open Output Folder', 'Delete Selected', 'Delete All', 'Print Selected Experiment (full)', and 'Print Selected Experiment (brief)'. The main area is divided into 'Experiment Sessions' on the left and 'Statistics' on the right. The 'Experiment Sessions' table lists various sessions with checkboxes for selection. The 'Statistics' section shows data for 'Experiment PPI20230626_155826', including a table of 'All cubicles totals' and a 'PERCENTAGE OF PPI' of 100. A graph on the right shows 'Pulse Alone' with a green bar at approximately 55%. At the bottom, a 'Stimuli [V]' graph shows a pulse at 0.25s with a duration of 100ms. The bottom status bar displays 'Amplitude mV: 3.765 Duration ms: 26 Latency s: 0.037' and the date '18/12/2023 15:45:49'.

9.1. Experiment sessions

On the left side, in the “experiment sessions” column, you can select the experiments to be included in the data grid. You can select a single session by clicking its name, or more session using Ctrl-Click or Shift-Click to add a single session or a group of them.

You can also double click a session name if you want to rename it:

The screenshot shows a 'Rename Session' dialog box. It contains a clipboard icon, the text 'Rename PPI_20231024_115026 to:', and a text input field containing 'PPI_20231024_115026'. At the bottom, there are 'Ok' and 'Cancel' buttons.

9.2. Data grid area

In the central area, the data grid contains the list of all the fired stimuli of the selected sessions.

The background of the lines is brownish when there has been a detection.

The columns have a drop-down that allows you to filter the data by one or more columns.

9.3. Plot chart area

When you select a stimulus line in the data grid, the relative curve is shown under the data grid:

Statistics Experiment 'PPI20230626_155826' Info

All cubicles totals	AVG Resp.%	Std.Dev.	Std.Err.	Q.ty	Det.
Pulse Alone	53.3	15.42	7.71	10	4
Prepulse-Pulse	0	0	0	0	0
PrePulse Alone	0	0	0	0	0
No Stimulus	0	0	0	0	0

PERCENTAGE OF PPI: 100

Enabled	#	Cubicle	Timer	StepName	Stimulus	Session	Protocol	Experiment
<input checked="" type="checkbox"/>	1	1	60.341		Pulse Alone	PPI_20230626_1558...	Rat Habituation.ppi	
<input checked="" type="checkbox"/>	1	2	60.341		Pulse Alone	PPI_20230626_1558...	Rat Habituation.ppi	
<input checked="" type="checkbox"/>	2	1	121.101		Pulse Alone	PPI_20230626_1558...	Rat Habituation.ppi	

Upper Threshold: 3.0 Lower Threshold: 1.5 Det. Timeout: 0.50

Amplitude mV: 3.765 Duration ms: 26 Latency s: 0.037

Here you can see the recorded period where:

- A green block indicates a pre-pulse
- A red block indicates a pulse
- A grey block indicates a no-stimulus
- A grayed area indicates the detection area

9.4. Stimuli exclusion


In certain circumstances you can decide that a stimulus is wrongly detected, as the reaction was not a real startle but due to other reasons.

In the central grid, the first column headed “Enabled” allows you to toggle the inclusion of a stimulus from the calculations. To exclude (or reinclude) a stimulus, just double-click its “Enabled” cell. The entire line is then grayed out to make clear the fact that it is excluded from the calculations.

Enabled	#	Cubicle	Timer	StepName	Stimulus	Session
<input checked="" type="checkbox"/>	1	1	60.341		Pulse Alone	PPI_20230621
<input type="checkbox"/>	1	2	60.341		Pulse Alone	PPI_20230621
<input checked="" type="checkbox"/>	2	1	121.191		Pulse Alone	PPI_20230621
<input checked="" type="checkbox"/>	2	2	121.191		Pulse Alone	PPI_20230621
<input checked="" type="checkbox"/>	3	1	182.051		Pulse Alone	PPI_20230621
<input checked="" type="checkbox"/>	3	2	182.051		Pulse Alone	PPI_20230621

9.5. Threshold adjustment

In the previous mentioned curve, you can change the thresholds if they were not good for a specific stimulus. Increasing it determines a higher sensitivity but also the risk of false positives. The opposite is true if we decrease it.



Upper Threshold Lower Threshold Det. Timeout [Save only this](#) [Save All same Session+Cubicle](#)
Amplitude mV: 3.039 Duration ms: 35 Latency s: 0.173

If you change them, a “Save only this” and a “Save All same Session+Cubicle” buttons appear.

Clicking “Save only this”, the new thresholds and detection timeout are saved to the database only for the selected stimulus.

Clicking “Save All same Session+Cubicle” the new thresholds and detection timeout are applied to all the stimuli for the current selected record's session and cubicle.

9.6. Exporting data

If you want to export to excel only selected sessions, select them in the left area and press “Export selected” button. Otherwise, all the sessions are saved.

For each session both a csv and a xlsx files are created. The xlsx file contains additional sheets: the second contains the raw data and the last contains the startle data.

Clicking the “Select fields to Export” button in the top menu, you can choose which fields you want to be included in the CSV/XLSX file.

Select Exported Fields

Select All Deselect All Please note that some fields are mandatory and cannot be deselected.

<input checked="" type="checkbox"/> Number	<input type="checkbox"/> Duration	<input type="checkbox"/> PrePulseWNIIntensity	<input type="checkbox"/> AnimalWeight
<input checked="" type="checkbox"/> Cubicle	<input checked="" type="checkbox"/> Latency	<input type="checkbox"/> PrePulseSoundIntensity	<input type="checkbox"/> Enabled
<input checked="" type="checkbox"/> Timer	<input type="checkbox"/> EnvFanActive	<input type="checkbox"/> PrePulseSoundFrequency	<input type="checkbox"/> PeakToPeak
<input type="checkbox"/> StepName	<input type="checkbox"/> EnvIRLightActive	<input type="checkbox"/> IPI	<input type="checkbox"/> DetectionTimeOut
<input checked="" type="checkbox"/> Stimulus	<input type="checkbox"/> EnvIRLightIntensity	<input type="checkbox"/> PulseDuration	
<input checked="" type="checkbox"/> SessionId	<input type="checkbox"/> EnvVisLightActive	<input type="checkbox"/> PulseWNI	
<input checked="" type="checkbox"/> ProtocolName	<input type="checkbox"/> EnvVisLightIntensity	<input type="checkbox"/> PulseSound	
<input checked="" type="checkbox"/> ExperimentName	<input type="checkbox"/> EnvWhiteNoiseActive	<input type="checkbox"/> PulseLight	
<input type="checkbox"/> TreatmentId	<input type="checkbox"/> EnvWNIIntensity	<input type="checkbox"/> PulseAirPuff	
<input type="checkbox"/> Stageld	<input type="checkbox"/> EnvAuxActive	<input type="checkbox"/> PulseOut1	
<input type="checkbox"/> Trial	<input type="checkbox"/> PreRecording	<input type="checkbox"/> PulseWNIIntensity	
<input checked="" type="checkbox"/> AnimalId	<input type="checkbox"/> PrePulseDuration	<input type="checkbox"/> PulseSoundIntensity	
<input type="checkbox"/> AnimalGender	<input type="checkbox"/> PrePulseWNI	<input type="checkbox"/> PulseSoundFrequency	
<input type="checkbox"/> AnimalColor	<input type="checkbox"/> PrePulseSound	<input type="checkbox"/> Thr_down	
<input checked="" type="checkbox"/> Detection	<input type="checkbox"/> PrePulseLight	<input type="checkbox"/> Thr_up	
<input checked="" type="checkbox"/> Amplitude	<input type="checkbox"/> PrePulseAirPuff	<input type="checkbox"/> RawData	
<input checked="" type="checkbox"/> StartlePercent	<input type="checkbox"/> PrePulseOut1	<input type="checkbox"/> StartleData	

You can click “Select All” or “Deselect All”. Please be aware that some fields are mandatory and cannot be excluded as they are needed to recognize uniquely a stimulus or contain the startle measures.

9.6.1. Raw data (XLSX only)

It's a value between 0 and 255, where 128 is the midline (2.5V) and 255 is 5V. A value every 1ms is given.

9.6.2. Startle data (XLSX only)

A value every 1ms is given. Only values different than 0 are significant:

- 1 means pre-pulse
- 4 means pulse
- 5 means no-stimulus
- If the value is greater than 15, it is the detection start point.

9.7. Statistical data


Selecting the “Statistics” tab on top of the data grid, a brief summary is shown:

Statistics Experiment 'PPI20231129_133343 lower sens' Info

All cubicles totals	AVG Resp.%	Std.Dev.	Std.Err.	Q.ty	Det.
Pulse Alone	56.8	18.36	4.33	20	18
Prepulse-Pulse	45.6	16.64	3.14	30	28
PrePulse Alone	22.4	0	0	30	1
No Stimulus	26.3	0	0	10	1

PERCENTAGE OF PPI

19.7




Enabled	#	Cubicle	Timer	StepName	Stimulus	Session	Protocol	Experiment
<input checked="" type="checkbox"/>	7	1	472.539	Phase B	Pulse Alone	PPI_20231129_1333...	BIO Protocoloni	Tonini

Here, the average, standard error, and standard deviation of the stimuli responses are shown. Only the stimuli with a detection are considered. The percentage of PPI at the bottom is the relationship between the pulses alone and the pulses preceded by a Pre-Pulse. “Pre-Pulse alone” and “No Stimulus” stimuli are not taken in account in this calculation.

The statistics include all the currently selected sessions.

10. Control Page

This page is used to test the various environment and stimuli in the cubicle. We suggest a quick check to this page before the tests.



The environment outputs are executed immediately.


The Startle outputs are fired for two second when you press the “FIRE STIMULUS” button.

This is just for testing purposes.

The two Aux outputs can be used to connect external devices. Out 1 is tied to the stimulus and the Out 2 is environmental.

If you click the cubicle round button in the CONNECTED CUBICLES area, the Startle PPI Cage light flashes twice and a low-volume sound is fired. This can help you to identify the cages as the numbering of the cubicles may depend on the order in which they are connected to the USB ports.

CONNECTED CUBICLES



11. Settings Page



11.1. Working Database

In this page you can select the database to be used to store the recorded data.


Be careful using this feature as you cannot view or compare data from different databases.

Here you can select a new database (the default one is called "db.sqlite"). You can choose to switch to a previously used database or to create a new database.

If you want to create a new database, its name must follow the file naming rules, so the following characters are not allowed: < > / \ | * ? : " .

Moreover, the database MUST reside in the default folder (in the "Documents\StartlePPI\DB" folder"). If you try to open a database that resides outside this folder, a warning appears. Please copy the needed database in that folder and then you can open it.

Please note that the currently selected database name is always visible in the application title bar (enclosed in squared brackets):



11.2. Firmware Update

This functionality is available starting from the app version V2.0.21.0.

This functionality allows you to update the firmware of the first connected cubicle.

The update file must be placed in the Documents\StartlePPI\Updates folder and its name has a ".s19" file extension.

It is suggested, when using this function, to connect only one cubicle at a time.

CUBICLE FIRMWARE UPDATE

Here you can update the first cubicle's firmware. IMPORTANT: Do not disconnect or power off the device while updating. Failure to update may result in the device failing to boot-up.



First of all, you must select the firmware update file pressing the “Select Firmware” button.

The name of the selected file is shown near the update button.

Then you can start the update process pressing the “Firmware Update” button:

CUBICLE FIRMWARE UPDATE

Here you can update the first cubicle's firmware. IMPORTANT: Do not disconnect or power off the device while updating. Failure to update may result in the device failing to boot-up.



If the current firmware is V7.26 or earlier, you will be asked to disconnect, and then reconnect, the cubicle's board power supply.

To be precise, you must disconnect both the power supply and the USB connectors, wait a few seconds and plug the connectors back in the sockets.


If the current firmware is V7.27 or greater, the reset process is automatic so you won't need to unplug/plug back in any cable.

The update process should complete within a couple of minutes.

Please give a look to the update log for errors or warnings. If you encounter any problem, please contact the Ugo Basile service team describing the error message you received.

WARNING: the update process will completely erase and rewrite the internal memory of the device. Any power loss or disconnection once the update process has been started, may result in the device failing to boot up and the need to be serviced by the Ugo Basile support team. So please ensure that the power supply is reliable and the cables securely plugged in before starting the update process.

Once the update completes, please switch to the “RUN” page and verify that the device reports the new firmware version:



If you have multiple cages, please disconnect the just-updated one and connect the next one. Then proceed in the same way until you have updated all the cages.

12. Maintenance

While any service of the instrument ought to be carried out by Ugo Basile personnel or by qualified personnel authorized by UGO BASILE organization, this manual section describes normal maintenance procedures which can be carried out at your facility.

UNPLUG THE MAIN CORD BEFORE CARRYING OUT ANY MAINTENANCE JOB

12.1. Cleaning/disinfection

The Startle PPI does not require any maintenance apart from normal cleaning.

Do not use organic solutions. Cotton wool and water can be used for cleaning purposes. For disinfection, use a non-alcoholic disinfectant, or H₂O₂.

12.2. Long Inactivity

The instrument does not require any particular maintenance after long inactivity, except cleaning.

12.3. Customer Support

For any further information you may desire concerning the use and/or maintenance of the Startle PPI, please do not hesitate to contact our service department (or our local distributor) either directly or via our support web form.

Customer Support Contacts

Phone: +39 0332 744574

E-mail: service@ugobasile.com

Web form: <https://ugobasile.com/support/support-request>

Before sending the instrument to our factory for repair, please contact our logistics department to obtain a return authorization number (RMA) and shipping/packing instructions. We may not be held responsible for damages during transport due to poor packing; whenever possible, please use the original packing.

13. Specification

Operation

Sound Range	From 65 dB to 120 dB
Pulse/pre pulse sound range	From 65 dB to 120 dB
Sound Frequency Range	From 100 Hz to 18 KHz
IR Light	0-100% (Environment)
Flash Light	0-100% (20K Lux) S/R Platform)
Data Export	.csv, .xlsx format

General

Power Requirements	Universal input 100-240 VAC, 50-60Hz, 60W
Background Noise (Ventilation)	from 0.1 to 3mA, in 0.1mA steps
Operating environment	10°C to 40°C; 5% to 95% RH (non-condensing)

Physical

Cubicle dimension, outside	52(d)x62(w)x58(h)cm
Cubicle dimension, inside	39(d)x47(w)x49(h)cm
Shipping Weight	Mouse Single-Cage Set-Up: 40Kg
Packing Dimensions	77x65x83 cm (wooden crate) for a Mouse Single-Cage Set-up

Warranty

Warranty	Startle PPI is covered by a 12-month warranty + 12 after product registration
----------	---


14. Bibliografy

- G. Corli, M. Tirri, S. Bilel, R. Arfè, T. Coccini, E. Roda, B. Marchetti, F. Vincenzi, G. Zauli, P. A. Borea, C. A. Locatelli, K. Varani, M. Marti, 2023, "[MAM-2201 acute administration impairs motor, sensorimotor, prepulse inhibition, and memory functions in mice: a comparison with its analogue AM-2201](#)", Psychopharmacology
- M. Tirri, S. Bilel, R. Arfè, G. Corli, B. Marchetti, T. Bernardi, F. Boccuto, G. Serpelloni, F. Botrè, F. De-Giorgio, K. Golembiowska, M. Marti, 2022, "[Effect of -NBOMe Compounds on Sensorimotor, Motor, and Prepulse Inhibition Responses in Mice in Comparison With the 2C Analogs and Lysergic Acid Diethylamide: From Preclinical Evidence to Forensic Implication in Driving Under the Influence of Drugs](#)", Frontiers in Psychiatry
- M. Marti, G. Talani, C. Miliano, S. Bilel, F. Biggio, J. Bratzu, M. Diana, M. A. De Luca, L. Fattore, 2021 "[New insights into methoxetamine mechanisms of action: Focus on serotonergic 5-HT2 receptors in pharmacological and behavioral effects in the rat](#)", Experimental Neurology
- M. Tirri, L. Ponzoni, S. Bilel, Raffaella Arfè, D. Braidà, M. Sala, M. Marti, 2020, "[Acute DOB and PMA Administration Impairs Motor and Sensorimotor Responses in Mice and Causes Hallucinogenic Effects in Adult Zebrafish](#)", Brain Sciences
- S. Bilel, M. Tirri, R. Arfè, S. Stopponi, L. Soverchia, R. Ciccocioppo, P. Frisoni, S. Strano-Rossi, C. Miliano, F. De-Giorgio, G. Serpelloni, A. Fantinati, A. A. De Luca, M. Neri, M. Marti, 2019, "[Pharmacological and Behavioral Effects of the Synthetic Cannabinoid AKB48 in Rats](#)", Frontiers in Neuroscience
- David L. Braff, Mark A. Geyer, Neal R. Swerdlow, 2001, "[Human studies of prepulse inhibition of startle: normal subjects, patient groups, and pharmacological studies](#)", Psychopharmacology
- M.Koch, 1999, "[The Neurobiology of Startle](#)" Progress in Neurobiology
- R. R. Marsh, H. S. Hoffman, C. L. Stitt, G. M. Schwartz, 1975, "[The Role of Small Changes in the Acoustic Environment in Modifying the Startle Reflex](#)" J Exp Psychol Anim Behav Process
- H.S. Hoffman, M. Fleshker, 1963, "[Startle Reaction: Modification By Background Acoustic Stimulation](#)", Science

15. Related products

Since more than 50 years Ugo Basile designs and manufactures classic and innovative devices, essential in animal behavioral studies.

Whether your research involves the study of memory, learning, anxiety, depression, fear, stress, social interaction, addiction, or more complex behavioural protocols, we have the answer: our line of conditioning/avoidance cages, mazes, videotracking, etc. keeps growing thanks to always new devices and methods.



46000 series- Fear Conditioning System



46000-101 - Electric Shock Meter



49603 - Operant Fear Conditioning



49500/49503 - Operon (for attentional set-shifting)



42552/42553 - Conditioned Place Preference Cages for Rats or Mice



47502/47503 Learned Helplessness



40532/40533 - Active Avoidance set-up for Mice and Rats (Shuttle-Box)



40570 - Passive Avoidance - Step Down for Mice (vibrating platform)



40550 - Passive Avoidance - Step through - for Mice and Rats



45100/40150 - Lickometer (Vogel Test) set-up for Drinking-Conflict experiments



Your Companion in Discovery since 1963

ugobasile.com

more than 40,000 citations in the main bibliographic search engines.



Ugo Basile SRL Società Unipersonale

Via Giuseppe Di Vittorio, 2

21036 Gemonio (VA) ITALY

Tel. +39 0332 744574

sales@ugobasile.com

service@ugobasile.com

